

# Sensium

## Declaration of Conformity

**Sensium Healthcare Ltd**  
**115 Olympic Avenue, Milton Park, Abingdon, OX14 4SA, UK**

Sensium Healthcare Ltd declare under our sole responsibility that

**Product: SensiumVitals Home Bridge**  
**Model Number SH20222x**

as part of the TZ202001 Sensium System, is in conformity with the relevant UK legislation:

S.I. 2017/1206 Radio Equipment Regulations 2017

S.I. 2012/3032 Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012

ETSI EN 300 220-2 V3.1.1 (2017-02)

ETSI EN 301 511 V12.5.1

ETSI EN 301 908-13 V13.1.1

ETSI EN 301 489-1 V2.2.0 (2017-03)

ETSI EN 301 489-3 V2.1.1 (2017-03)

ETSI EN 301 489-52 V1.1.0 (2016-11)

EN 60601-1-2:2015 + A1:2014

EN 60601-1:2006+A1:2013+A12:2014

EN 60601-1-11:2015 + A1:2021

EN IEC 62311:2020

EN IEC 63000:2018

EN 50419:2006

Signature



Name Bruce Slaymaker

Position Regulatory and QA Manager

Date 31<sup>st</sup>-July-2021