

## Declaration of Conformity

**Sensium Healthcare Ltd**  
**Building 3, 115 Milton Park, Abingdon, Oxon, OX14 4SA, UK**

Sensium Healthcare Ltd declares that the following product is in conformity with the Essential Requirements and provisions of the following Directives and standards:

99/5/EEC	RTTE Directive
2011/62/EU	RoHS2 Directive
2012/19/EU	WEEE Directive
EN 60950-1:2006 + A1:2010	Information Technology Equipment - Safety
ETSI EN 300 220-2 V2.4.4 (2012-05)	Short Range Devices 25 to 1000 MHz.
ETSI EN 300 328 V1.8.1 (2012-06)	Wideband transmission systems; in the 2.4 GHz ISM band.
ETSI EN 301 489-1 V1.9.2 (2011-09)	EMC for Radio Equipment – Common Technical Requirements
ETSI EN 301 489-3 V1.6.1 (2013-08)	EMC for Short Range Devices 9 kHz to 246 GHz.
ETSI EN 301 489-17 V2.2.1 (2012-09)	EMC for Broadband Data Transmission Systems
EN 62479:2010	Human exposure to electromagnetic fields

**Product :** SensiumVitals® Bridge  
**Model Number** TZ202022

**Signature**



**Name** Bruce Slaymaker  
**Position** Regulatory and QA Manager

**Date** 22<sup>nd</sup> Jan 2016

**CE**  
**0889**